

EU DECLARATION OF CONFORMITY FOR MEDICAL DEVICES

Name and Address of the Manufacturer: Vertigocatcher Diagnostics AB

SRN of the Manufacturer: TBD

Basic UDI-DI: 735016159NYTL8

Product Description: The Nystagmocatcher hardware is mounted on the camera lens of a cellphone. The hardware consists of a lens, a lamp and a clamp to mount the hardware on the cellphone. The hardware has a soft foam that is hold tight around the eye by holding the cellphone with one hand.

Product Identification: N/A

Product-No.: N/A

Classification (acc. to Annex VIII of MDR): Class I


Conformity Assessment Procedure: Self-declaration acc. to Annex I,II, III and IV Regulation (EU) 2017/745 on Medical Devices

On our sole responsibility, we hereby declare that the product(s) comply with the relevant provisions of the following Regulation and Directives:

Regulation (EU) MDR 2017/745 of the European parliament and of the Council on Medical Devices

This declaration is valid until the next change of the product. Any modification of the medical device not authorized by us will invalidate this declaration.

Signature:



Christian Geisler

Date:

2024-04-16 Stockholm

Place:

**ANNEX I
TO THE EU DECLARATION OF CONFORMITY
FOR MEDICAL DEVICES**

Products Covered:

| Trade Name | Article No. |
|-------------------|--------------------|
| Nystagmocatcher | 0124 |